

1 **Spinal Plating Systems – Performance**
2 **Criteria for Safety and Performance**
3 **Based Pathway**

6 **Draft Guidance for Industry and**
7 **Food and Drug Administration Staff**

9 ***DRAFT GUIDANCE***

11 **This draft guidance document is being distributed for comment purposes**
12 **only.**

14 **Document issued on September 20, 2019.**

16 You should submit comments and suggestions regarding this draft document within 90 days of
17 publication in the *Federal Register* of the notice announcing the availability of the draft
18 guidance. Submit electronic comments to <https://www.regulations.gov>. Submit written
19 comments to the Dockets Management Staff (HFA-305), Food and Drug Administration, 5630
20 Fishers Lane, rm. 1061, Rockville, MD 20852. Identify all comments with the docket number
21 listed in the notice of availability that publishes in the *Federal Register*.

23 For questions about this document, contact the DHT6B: Division of Spinal Devices at 301-796-
24 5650 or Jonathan Peck at Jonathan.Peck@fda.hhs.gov.



28 **U.S. Department of Health and Human Services**
29 **Food and Drug Administration**
Center for Devices and Radiological Health

Preface

31
32
33
34
35
36
37
38

Additional Copies

Additional copies are available from the Internet. You may also send an e-mail request to CDRH-Guidance@fda.hhs.gov to receive a copy of the guidance. Please include the document number 19008 and complete title of the guidance in the request.

DRAFT

Spinal Plating Systems – Performance Criteria for Safety and Performance Based Pathway

Draft Guidance for Industry and Food and Drug Administration Staff

This draft guidance, when finalized, will represent the current thinking of the Food and Drug Administration (FDA or Agency) on this topic. It does not establish any rights for any person and is not binding on FDA or the public. You can use an alternative approach if it satisfies the requirements of the applicable statutes and regulations. To discuss an alternative approach, contact the FDA staff or Office responsible for this guidance as listed on the title page.

I. Introduction

This draft guidance provides performance criteria for spinal plating systems in support of the [Safety and Performance Based Pathway](#).¹ Under this framework, submitters planning to submit a 510(k) using the Safety and Performance Based Pathway for spinal plating systems will have the option to use the performance criteria proposed in this draft guidance to support substantial equivalence, rather than a direct comparison of the performance of the subject device to that of a predicate device.

For the current edition of the FDA-recognized standard(s) referenced in this document, see the [FDA Recognized Consensus Standards Database](#).² For more information regarding use of consensus standards in regulatory submissions, please refer to the FDA guidance titled [Appropriate Use of Voluntary Consensus Standards in Premarket Submissions for Medical Devices](#).³

FDA's guidance documents, including this draft guidance, do not establish legally enforceable responsibilities. Instead, guidances describe the Agency's current thinking on a topic and should

¹ Available at <https://www.fda.gov/regulatory-information/search-fda-guidance-documents/safety-and-performance-based-pathway>

² Available at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfStandards/search.cfm>

³ Available at <https://www.fda.gov/regulatory-information/search-fda-guidance-documents/appropriate-use-voluntary-consensus-standards-premarket-submissions-medical-devices>

Contains Nonbinding Recommendations

Draft – Not for Implementation

70 be viewed only as recommendations, unless specific regulatory or statutory requirements are
71 cited. The use of the word *should* in Agency guidance means that something is suggested or
72 recommended, but not required.
73

74 **II. Scope/Device Description**

75 The spinal plates that are the subject of this guidance are anterior cervical or anterior/lateral
76 thoracolumbar spinal plating systems. These devices are Class II and are regulated under 21 CFR
77 888.3060 with the product code KWQ (appliance, fixation, spinal intervertebral body). General
78 guidance on submission of a 510(k) for a spinal plating system can be found in FDA’s guidance
79 [Spinal System 510\(k\)s](#).⁴
80

81 **Intended Use/Indications for Use:** The spinal plating systems that fall within the scope of this
82 guidance document are intended for fixation to vertebral bodies (anteriorly in the cervical spine
83 or anteriorly/laterally in the thoracolumbar spine) for the purpose of stabilizing the spine for
84 fusion. Plating systems that attach to the posterior spine are outside the scope of this guidance
85 document.
86

87 **Device Design Characteristics:** The spinal plating systems that fall within the scope of this
88 guidance document consist of plates and associated fixed or variable angle screws, constructed
89 solely from one of the following titanium alloys in conformance with the associated FDA-
90 recognized consensus standard:

- 91 • American Society for Testing and Materials (ASTM) F136 *Standard Specification for*
92 *Wrought Titanium-6 Aluminum-4 Vanadium ELI (Extra Low Interstitial) Alloy for*
93 *Surgical Implant Applications (UNS R56401)*
- 94 • ASTM F1295 *Standard Specification for Wrought Titanium-6 Aluminum-7Niobium Alloy*
95 *for Surgical Implant Applications (UNS R56700)*
- 96 • ASTM F67 *Standard Specification for Unalloyed Titanium, for Surgical Implant*
97 *Applications (UNS R50250, UNS R50400, UNS R50550, UNS R50700).*
98

99 A dimensional comparison of the subject device should be performed, and the dimensions should
100 fall within the dimensional ranges listed in Table 1.
101

⁴ Available at <https://www.fda.gov/regulatory-information/search-fda-guidance-documents/guidance-industry-and-fda-staff-spinal-system-510ks>

Contains Nonbinding Recommendations

Draft – Not for Implementation

102 **Table 1** - Size ranges for cervical and thoracolumbar spinal plating systems.

Cervical Plates	Range*
Number of Levels Treated	1 to 5
Plate Length (hole-to-hole)	10 mm to 115 mm
Plate Thickness/Profile**	≤ 3 mm
Screw Diameter (Major)	3.5 mm to 4.5 mm
Screw Length (Threaded Length)	10 mm to 26 mm
Thoracolumbar Plates	
Number of Levels Treated	1 to 3
Plate Length (hole-to-hole)	15 mm to 130 mm
Plate Thickness/Profile**	≤ 7 mm
Screw Diameter (Major)	5 mm to 7 mm
Screw Length (Threaded Length)	15 mm to 70 mm

103
104 * The dimensional ranges listed were derived from historical data submitted to FDA in 510(k) submissions for
105 devices previously found substantially equivalent.

106 ** Largest thickness or profile of the subject plate should fall below the listed value.

107
108 Cervical and thoracolumbar spinal plating systems with the following features are not eligible
109 for the Safety and Performance Based Pathway via this guidance:

- 110 • Devices that affix to the posterior spine
- 111 • Devices for which a 2-level cervical plate or a 1- or 2-level thoracolumbar plate is not
112 representative of a worst-case construct for performance testing per the FDA currently
113 recognized version of ASTM F1717 *Standard Test Methods for Spinal Implant*
114 *Constructs in a Vertebrectomy Model*
- 115 • Staples or plates with fixation mechanisms other than threaded screws
- 116 • Devices with coatings
- 117 • Combination products
- 118 • Resorbable devices
- 119 • Additively manufactured devices
- 120 • Devices that are designed to allow motion post-implantation (e.g., plates designed to
121 “settle”).
- 122 • Buttress plating systems (i.e., plates that do not span at least one functional spinal unit)

123
124 Where FDA determines that additional data are necessary to make these determinations, the
125 Agency may, on a case-by-case basis, review that data before determining whether or not the
126 device is appropriate for the Safety and Performance Based Pathway. In situations, where you
127 determine that additional testing outside of those identified in this guidance are necessary to
128 make a determination regarding eligibility into the Safety and Performance Based Pathway, we

Contains Nonbinding Recommendations

Draft – Not for Implementation

129 would encourage sponsors to submit a Pre-Submission⁵ to engage in discussion with FDA prior
130 to submission of the 510(k).
131

132 **III. Testing Performance Criteria**

133 If your device is appropriate for submission through the Safety and Performance Based Pathway,
134 and you choose to use that option, you do not need to provide direct comparison testing against a
135 legally marketed predicate to demonstrate substantially equivalent performance characteristics.
136 To ensure that the performance criteria outlined in this guidance remain contemporary and take
137 into account relevant data from recent clearances, FDA recommends that you provide a results
138 summary for all tests evaluated in addition to the other submission information (e.g., Declaration
139 of Conformity (DoC)) identified for each test or evaluation below. Unless otherwise identified in
140 the submission information sections below, test information such as results summary, test
141 protocols, or complete test reports should be submitted as part of the 510(k) as described in
142 FDA’s guidance, [Safety and Performance Based Pathway](#).⁶ For additional information regarding
143 the submission of non-clinical bench testing information, please see FDA’s guidance
144 [Recommended Content and Format of Non-Clinical Bench Performance Testing Information in
145 Premarket Submissions](#).⁷
146

147 **Mechanical Testing**

148
149 Static compression bending, static torsion, and dynamic compression bending should be
150 performed in conformance with the FDA currently-recognized version of ASTM F1717
151 *Standard Test Methods for Spinal Implant Constructs in a Vertebrectomy Model*. We recommend
152 that you perform all testing on plate system designs that represent worst-case (e.g., most likely to
153 loosen or fail) final design versions. You should also provide a rationale identifying how you
154 identified the worst-case design. Acceptance criteria are listed below for each test, which
155 include stiffness and yield values for the static tests and runout loads for the dynamic test.⁸
156

157 For each mechanical test below, you should provide a report as specified in the relevant reporting
158 sections of ASTM F1717 and the Mechanical Testing section of FDA’s guidance [Spinal System
159 510\(k\)s](#),⁹ in addition to a Declaration of Conformity (DoC) to the consensus standard. Any

⁵ Available at <https://www.fda.gov/regulatory-information/search-fda-guidance-documents/requests-feedback-and-meetings-medical-device-submissions-q-submission-program>

⁶ Available at <https://www.fda.gov/regulatory-information/search-fda-guidance-documents/safety-and-performance-based-pathway>

⁷ Available at <https://www.fda.gov/regulatory-information/search-fda-guidance-documents/recommended-content-and-format-non-clinical-bench-performance-testing-information-premarket>

⁸ It should be noted that although ASTM F1717 is FDA-recognized in full, FDA believes that for the purposes of the safety and performance based pathway, the testing, methods and criteria identified in this section on mechanical bench testing represent the least burdensome approach to demonstrating substantial equivalence for this pathway, although alternative or additional methods or acceptance criteria are identified in the recognized consensus standard for some tests.

⁹ Available at <https://www.fda.gov/regulatory-information/search-fda-guidance-documents/guidance-industry-and-fda-staff-spinal-system-510ks>

Contains Nonbinding Recommendations

Draft – Not for Implementation

160 protocol deviations should be thoroughly described and justified; however, note that certain
161 protocol deviations may invalidate comparison to the performance criteria listed below, resulting
162 in the need for submission of a Traditional, Special, or Abbreviated 510(k), as appropriate.

163
164 **Note:** ASTM F1717 specifies the active lengths of the longitudinal element to be 35 mm for
165 cervical devices and 76 mm for lumbar devices (or as close to these dimensions as possible based
166 on plate sizes available) to simulate connection across two spinal levels in the cervical and
167 lumbar spine, respectively. However, since many thoracolumbar plating systems only contain 1-
168 level plates, significant modification to the specified 76 mm active length is necessary to
169 simulate connection across a single spinal level. Therefore, data for 1-level and 2-level
170 thoracolumbar plating systems were analyzed separately, and acceptance criteria are stratified for
171 each test below.

- 172
173 1. **Test name:** ASTM F1717 - Static compression bending
174 **Methodology:** ASTM F1717 *Standard Test Methods for Spinal Implant Constructs in a*
175 *Vertebrectomy Model*
176 **Performance Criteria:**

177
178 **Table 2** –Static compression bending acceptance criteria for cervical and thoracolumbar
179 plating systems

Test Parameter	Cervical (2-Level constructs)	Thoracolumbar (1-level constructs)	Thoracolumbar (2-level constructs)
Static Compression Bending Stiffness (N/mm)	9.6 N/mm	45 N/mm	35 N/mm
Static Compression Bending Yield (N)	75 N	230 N	360 N

180
181
182 **Performance Criteria Source:** Criteria are based on aggregated mechanical testing data
183 submitted to FDA in 510(k) submissions for spinal plating systems previously found to
184 be substantially equivalent.

185 **Additional Considerations:** Testing should include a minimum of 5 samples consistent
186 with ASTM F1717. In order to be considered a successful result, either: (1) all samples
187 should meet or exceed the acceptance criteria listed above, or (2) the average of all
188 samples should meet or exceed the criteria above and the standard deviation should be \leq
189 10% of the calculated average. For testing of 1-level thoracolumbar plates, active length
190 for the worst case should fall between 25 and 40 mm to be comparable to the criteria
191 listed in the table above.

192 **Submission Information:** Results summary and DoC

- 193
194 2. **Test name:** ASTM F1717 - Static torsion
195 **Methodology:** ASTM F1717 *Standard Test Methods for Spinal Implant Constructs in a*
196 *Vertebrectomy Model*

Contains Nonbinding Recommendations

Draft – Not for Implementation

Performance Criteria:

Table 3 – ASTM F1717 static torsion acceptance criteria for cervical and thoracolumbar plating systems.

Test Parameter	Cervical (2-Level constructs)	Thoracolumbar (1-level constructs)	Thoracolumbar (2-level constructs)
Static Torsion Stiffness (N-m/degree)	0.9 N-m/degree	5.6 N-m/degree	2.7 N-m/degree
Static Torsion Yield (N-m)	4.7 N-m	19 N-m	18 N-m

Performance Criteria Source: Criteria are based on aggregated mechanical testing data submitted to FDA in 510(k) submissions for spinal plating systems previously found to be substantially equivalent.

Additional Considerations: Testing should include a minimum of 5 samples consistent with ASTM F1717. In order to be considered a successful result, either: (1) all samples should meet or exceed the acceptance criteria listed above, or (2) the average of all samples should meet or exceed the criteria above and the standard deviation should be \leq 10% of the calculated average. For testing of 1-level thoracolumbar plates, active length for the worst case should fall between 25 and 40 mm to be comparable to the criteria listed in the table above.

Submission Information: Results summary and DoC

3. **Test name:** ASTM F1717 - Dynamic compression bending fatigue test

Methodology: ASTM F1717 *Standard Test Methods for Spinal Implant Constructs in a Vertebrectomy Model*

Performance Criteria:

Table 4 – ASTM F1717 dynamic compression bending acceptance criteria for cervical and thoracolumbar plating systems.

Test Parameter	Cervical (2-Level constructs)	Thoracolumbar (1-level constructs)	Thoracolumbar (2-level constructs)
Dynamic Compression Bending Runout Load to 5 Mc (N)	40 N	165 N	165 N

Performance Criteria Source: Criteria are based on aggregated mechanical testing data submitted to FDA in 510(k) submissions for spinal plating systems previously found to be substantially equivalent.

Contains Nonbinding Recommendations

Draft – Not for Implementation

227 **Additional Considerations:** Fatigue testing should include a minimum of 6 samples
228 with at least two runouts at the highest established runout load and at least one failure.
229 Fatigue precision (the ratio of the lowest failure load to the highest established runout)
230 should meet the level specified in ASTM F1717. For testing of 1-level thoracolumbar
231 plates, active length for the worst case should fall between 25 and 40 mm to be
232 comparable to the criteria listed in the table above.

233 **Submission Information:** Results summary and DoC
234

235 **Sterilization (devices labeled as sterile) and Reprocessing (end-user sterilized) Validation**
236

237 4. **Test name:** Sterilization (devices labeled as sterile) and Reprocessing (end-user
238 sterilized)

239 **Methodology:** FDA currently-recognized versions of the following consensus standards
240 (as applicable):

- 241 • International Organization for Standardization (ISO) 17665-1 *Sterilization of*
242 *health care products – Moist heat – Part 1: Requirements for the development,*
243 *validation, and routine control of a sterilization process for medical devices*
- 244 • ISO 11135-1 *Sterilization of health care products – Ethylene oxide- Part 1:*
245 *Requirements for development, validation, and routine control of a sterilization*
246 *process for medical devices*
- 247 • ISO 11137-1 *Sterilization of health care products—Radiation—Part 1:*
248 *Requirements for development, validation, and routine control of a sterilization*
249 *process for medical devices*
- 250 • ISO 11607-1 *Packaging for terminally sterilized medical devices – Part 1:*
251 *Requirements for materials, sterile barrier systems and packaging systems*
- 252 • ISO 11607-2 *Packaging for terminally sterilized medical devices – Part 2:*
253 *Validation requirements for forming, sealing and assembly processes*

254 **Performance Criteria:** Validation testing should demonstrate the cleanliness and
255 sterility of, or the ability to clean and sterilize to a sterility assurance level of 10^{-6} , the
256 device and device-specific instruments. You should provide a description of the
257 packaging (sterile barrier system) and how it will maintain the device’s sterility, and a
258 description of the package test methods, but not package test data.

259 **Performance Criteria Source:** FDA’s guidance:

- 260 • [Submission and Review of Sterility Information in Premarket Notification](#)
261 [\(510\(k\) Submissions for Devices Labeled as Sterile](#)¹⁰
- 262 • [Reprocessing Medical Devices in Health Care Settings: Validation Methods and](#)
263 [Labeling](#)¹¹

264 **Submission Information:** If using an Established Category A sterilization method, you
265 should provide the information described in Section V.A. as specified in the FDA
266 guidance [Submission and Review of Sterility Information in Premarket Notification](#)

¹⁰ Available at <https://www.fda.gov/regulatory-information/search-fda-guidance-documents/submission-and-review-sterility-information-premarket-notification-510k-submissions-devices-labeled>

¹¹ Available at <https://www.fda.gov/regulatory-information/search-fda-guidance-documents/reprocessing-medical-devices-health-care-settings-validation-methods-and-labeling>

Contains Nonbinding Recommendations

Draft – Not for Implementation

267 [\(510\(k\)\) Submissions for Devices Labeled as Sterile](#); the validation data itself is not
268 needed to demonstrate substantial equivalence.

269

270 **Biocompatibility Evaluation:**

271

272 To identify the biocompatibility endpoints to include as part of your biocompatibility evaluation
273 you should use Attachment A of CDRH’s guidance [Use of International Standard ISO 10993-1,](#)
274 [Biological evaluation of medical devices – Part 1: Evaluation and testing within a risk](#)
275 [management process](#),¹² referred to in the rest of this document as the “CDRH Biocompatibility
276 Guidance” for brevity. FDA considers the devices covered by this guidance to be categorized as
277 Implant Devices in contact with tissue/bone with a permanent contact duration of > 30 days and
278 you should assess the endpoints below per Attachment A of the CDRH Biocompatibility
279 Guidance.

- 280 • Cytotoxicity
- 281 • Sensitization
- 282 • Irritation or Intracutaneous Reactivity
- 283 • Acute Systemic Toxicity
- 284 • Material-Mediated Pyrogenicity
- 285 • Sub-acute/Sub-chronic Toxicity
- 286 • Genotoxicity
- 287 • Implantation
- 288 • Chronic Toxicity
- 289 • Carcinogenicity

290

291 **Rationale in Lieu of Testing:** If the subject device is manufactured from the identical raw
292 materials using identical manufacturing processes as a predicate device with the same type and
293 duration of tissue contact, and any changes in geometry are not expected to impact the biological
294 response, this is typically sufficient to establish substantially equivalent biocompatibility if
295 documentation such as that outlined in Attachment F of the CDRH Biocompatibility Guidance is
296 also provided.

297

298 **Testing:** In rare cases, if you determined that testing is needed to address some or all of the
299 identified biocompatibility endpoints, FDA recommends that complete test reports be provided
300 for all tests performed unless a declaration of conformity without supplemental information can
301 be appropriately provided, per Attachment E of the CDRH Biocompatibility Guidance. Any test-
302 specific positive, negative, and/or reagent controls should perform as expected, and protocol
303 deviations should be thoroughly described and justified; however, note that certain protocol
304 deviations may invalidate comparison to the performance criteria listed below, resulting in the
305 need for submission of a Traditional, Special, or Abbreviated 510(k).

306

307

¹² Available at <https://www.fda.gov/regulatory-information/search-fda-guidance-documents/use-international-standard-iso-10993-1-biological-evaluation-medical-devices-part-1-evaluation-and>

Contains Nonbinding Recommendations

Draft – Not for Implementation

- 308 5. **Test name:** Biocompatibility endpoints (identified from CDRH Biocompatibility
309 Guidance)
310 **Methodology:** FDA currently-recognized versions of biocompatibility consensus
311 standards
312 **Performance Criteria:** All direct or indirect tissue contacting components of the device
313 and device-specific instruments should be determined to have an acceptable biological
314 response.
315 **Performance Criteria Source:** The CDRH Biocompatibility Guidance
316 **Additional Considerations:** For any biocompatibility test samples with an adverse
317 biological response, the biocompatibility evaluation should explain why the level of
318 toxicity seen is acceptable. Some comparison testing against a legally marketed predicate
319 may be necessary (and is considered acceptable under the Safety and Performance Based
320 Pathway) to support such a rationale as explained in the CDRH Biocompatibility
321 Guidance. For standard biocompatibility test methods that include comparison device
322 control samples, the legally marketed comparison device control samples should perform
323 as expected, as specified above for the subject device samples.
324 **Submission Information:** Refer to CDRH Biocompatibility Guidance

DRAFT